UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/612,780	07/01/2003	Michael W. Dae	RADNT-039C	5891
Robert D. Buyan Stout, Uxa, Buyan & Mullins, LLP			EXAMINER	
			GIBSON, ROY DEAN	
4 Venture, Suite 300 Irvine, CA 92618			ART UNIT	PAPER NUMBER
,			3739	
SHORTENED STATUTOR	LY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE	
3 MONTHS		02/12/2007	PAPER	

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

	Application No.	Applicant(s)				
	10/612,780	DAE ET AL.				
Office Action Summary	Examiner	Art Unit				
	Roy D. Gibson	3739				
The MAILING DATE of this communication app Period for Reply	ears on the cover sheet with the c	orrespondence address				
A SHORTENED STATUTORY PERIOD FOR REPLY WHICHEVER IS LONGER, FROM THE MAILING DA - Extensions of time may be available under the provisions of 37 CFR 1.13 after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period w - Failure to reply within the set or extended period for reply will, by statute, Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICATION 16(a). In no event, however, may a reply be tin ill apply and will expire SIX (6) MONTHS from cause the application to become ABANDONE	N. nely filed the mailing date of this communication. D (35 U.S.C. § 133).				
Status	•					
1) Responsive to communication(s) filed on 22 No.	ovember 2006.					
,	action is non-final.					
closed in accordance with the practice under E						
Disposition of Claims						
4)⊠ Claim(s) <u>42-48 and 52-69</u> is/are pending in the application.						
4a) Of the above claim(s) is/are withdrawn from consideration.						
5)⊠ Claim(s) <u>53-69</u> is/are allowed.						
6)⊠ Claim(s) <u>42,45-48 and 52</u> is/are rejected.						
7)⊠ Claim(s) <u>43 and 45</u> is/are objected to.						
8) Claim(s) are subject to restriction and/or	election requirement.					
Application Papers						
9) The specification is objected to by the Examine	r.					
10) The drawing(s) filed on is/are: a) accepted or b) objected to by the Examiner.						
Applicant may not request that any objection to the	drawing(s) be held in abeyance. Se	e 37 CFR 1.85(a).				
Replacement drawing sheet(s) including the correct	ion is required if the drawing(s) is ob	jected to. See 37 CFR 1.121(d).				
11)☐ The oath or declaration is objected to by the Ex	aminer. Note the attached Office	Action or form PTO-152.				
Priority under 35 U.S.C. § 119						
12) ☐ Acknowledgment is made of a claim for foreign a) ☐ All b) ☐ Some * c) ☐ None of:	priority under 35 U.S.C. § 119(a)-(d) or (f).				
1. Certified copies of the priority documents have been received.						
2. Certified copies of the priority documents have been received in Application No						
3. Copies of the certified copies of the priority documents have been received in this National Stage						
application from the International Bureau	ı (PCT Rule 17.2(a)).					
* See the attached detailed Office action for a list of the certified copies not received.						
	-					
Attachment(s)						
1) Notice of References Cited (PTO-892)	4) Interview Summary					
2) Notice of Draftsperson's Patent Drawing Review (PTO-948)	Paper No(s)/Mail D 5) Notice of Informal F					
Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date	6) Other:	atom Approation				

Formal Matters

In December 2006 the examiner contacted Robert Buyan regarding an amendment of claim 42 which at the time he thought would place the application in Condition for Allowance. Subsequently the examiner's search resulted in an earlier disclosure of an anti-shivering agent being used with a heat exchange catheter as presented below.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 42, 45-48 and 52 are rejected under 35 U.S.C. 102(e) as being anticipated by Dobak, III (6,843,800). Dobak discloses a system comprising:

a heat exchange catheter (Figure 1) insertable into a blood vessel of the patient;

a temperature controller (not shown) in communication with the heat exchange catheter <u>inherently capable of</u> causing the heat exchange catheter to reduce the temperature of at least a patient's kidneys to a temperature at which a substance-induced renal damage is prevented or mitigated (a maximum temperature differential

Application/Control Number: 10/612,780

Art Unit: 3739

between the blood stream and the and the cooling element of the catheter to approximately 32 deg. C); and

an anti-shivering agent (meperidine which is a opioid receptor agonists) and col. 4, lines 12-43, col. 7, lines 12-18, col. 8, line 22-col. 9, line 28, col. 11, line 64-col. 12, line 39 and col. 13, line 63-col. 15, line 10).

Allowable Subject Matter

Claims 53-69 are allowed.

Claims 43 and 44 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Roy D. Gibson whose telephone number is 571-272-4767. The examiner can normally be reached on Tu-Th, 7:30 am-4:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda Dvorak can be reached on 571-272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/612,780

Art Unit: 3739

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Roy D. Gibson / Primary Examiner Art Unit 3739 Page 4

February 6, 2007